

# CARE PRODUCT AS



## EC DECLARATION OF CONFORMITY

Manufacturer: As above

Authorised Representative: As above  
(established within EU)

Details of electrical equipment

Model Number: 700

Description: Mattresspump for medical use.

Directives this equipment complies with: Relevant requirements in directive 93/42/EEC for medical equipment

Harmonised Standards applied in order to verify compliance with Directive: EMC standard: EN 60601-1-2  
Safety std.: EN 60601-1 (1990) + A1 (1993) + A2 (1995) + A12 (1993) +  
+ Corrigenda /July 1994)

Mattress 650MF and cushions

Dermatology at  
allergy is

The plastic folie has been tested by the Department of Enviromental  
Malmö University Hospital, Sweden and they conclude that the risk of  
negligible by skin contact.

Fire 037,  
equivalent.

The plastic folie conforms with the firetest according to the EN 597-1/NT  
Smouldering cigarette and the EN597-2/NT Fire 037, Match flame

Test Reports issued by: Notified Competent Body: NEMKO, Norway  
Report no: NEMKO reg. no: 199714151

Technical Construction File No: Not applicable

Year in wich CE mark was affixed: 1997

Authorised Signatory:

Manufacturer

Authorised Representative  
Established within the EU

Date of Issue  
21/12/1997

Name  
Position

Oddvar Kristiansen  
Managing Director

Oddvar Kristiansen  
Managing Director

### Head office

Bergkrystallen 25  
N - 2818 Gjøvik • Norway

Tel: +47 61 17 39 41  
Fax: +47 61 13 27 77

Mob: +47 90 03 64 49  
Mob.Fax: +47 90 27 64 49

### Sales office

Olav Kyrresgt. 6B  
N - 0273 Oslo • Norway

Tel: +47 22 12 33 90  
Fax: +47 22 55 56 49

Mob: +47 92 82 81 03  
Mob.Fax: +47 92 97 17 60