# Propad™ Visco Overlay



Dual layered visco elastic overlay



## Invacare Propad Visco Overlay

The **Propad Visco Overlay** is composed of a polyurethane cover and a high density visco elastic (memory) foam surface with a robust foam base.

This high specification surface is temperature sensitive allowing it to conform to the shape of the body and maximise contact area. This facilitates pressure redistribution and minimises pressure on key areas of the body such as the sacral and heels.


#### **Comfort and Conformity**

The interior of the overlay has a top layer of high density visco elastic foam, which is temperature-sensitive. The foam warms and conforms to body shape, maximising the body contact area. This provides exceptional comfort and facilitates pressure reduction by minimising peak pressures.

#### High density supportive foam base

The dual layer construction, combines a high level of comfort with improved longevity, due to the supportive, robust foam base.

# Features and Options





Cover is multi stretch, water-resistant and vapour-permeable.



### Available in a range of sizes

The *Invacare* **Propad Visco Overlay** is available in 2 standard sizes:
Single: 1875 x 876 x 85 mm,
Double: 1875 x 1372 x 85 mm.
Additional sizes available on request.



# Covers have an anti-slip base

Ensures safe use of the **Propad Visco Overlay** and helps prevent slippage.

#### Technical data

# For more information about this product, including the product's user manual, please visit www.invacare.co.uk



Size: 1875 x 876 mm <sup>1</sup> 1875 x 1372 mm <sup>2</sup> Additional sizes

available on request

85 mm

Height

1 2



7.1 kg



108 kg



Max. cleaning

80°



Violet Quartz



Manufactured to European Standards: BS EN 597-1 Cig, BS EN 597-2 Match and BS 7175 (Crib 5)

1. Single

Propad Visco

Overlav

2. Double

Essential nursing care is pivotal in pressure ulcer prevention. This product will positively contribute to the outcome of a pressure ulcer prevention care plan. Education, clinical judgement and action based planning based on vulnerability are fundamental factors in the prevention of pressure ulcers. A range of assessment scales can be used as a formal method of assessing risk from pressure ulcer development, and should be used in conjunction with an informal assessment (informed nursing judgement). Informal assessment is considered to be of greater importance and clinical value.

Invacare Limited
Pencoed Technology Park - Pencoed
Bridgend CF35 5AQ - United Kingdom
Tel: +44 1656 776 222
Fax: +44 1656 776 220
E-mail: uk@invacare.com
Sales Order E-mail:
ordersuk@invacare.com
www.invacare.co.uk

Invacare Ireland Ltd
Unit 5 - Seatown Business Campus
Seatown Road - SWORDS County Dublin - Ireland
Tel. +353 1 8107084
Fax +353 1 8107085
Email: Ireland@invacare.com
www.invacare.ie



