

Handled by, department  
Mikael Videby  
Building Technology and Mechanics  
+46 10 516 50 36, Mikael.Videby@sp.se

Invacare Rea AB  
Box 200  
343 75 Diö

## Crash testing of wheelchair XLT 75° according to ISO 7176-19:2001

(2 appendices)

### Summary

A crash test of the XLT 75° wheelchair has been performed according to ISO 7176-19:2001, chapter 5.2. The wheelchair was crash tested with a Hybrid III-dummy, 76 kg, in 48-50 km/h.

Standard Chapter		Comment	Fulfilment of requirement
5.2.1a	Horizontal excursion		Yes
5.2.1b	Knee vs. WC excursion		Yes
5.2.1c	Battery movement		N/A
5.2.2a	ATD torso angle <45°		Yes
5.2.2b	Securement points		Yes
5.2.2c	Separation of <100 grams		Yes
5.2.2d	Sharp edges radius > 2 mm		Yes
5.2.2e	Load carrying components		Yes
5.2.2f	Tilt-in space locking		N/A
5.2.2g	Removal of ATD, no tools		Yes
5.2.2h	WC removal, no tools		Yes
5.2.2i	Post height > 20%		Yes

The test object fulfilled the requirements according to ISO 7176-19:2001, chapter 5.2.

### SP Technical Research Institute of Sweden

Postal address  
SP  
Box 857  
SE-501 15 Borås  
SWEDEN

Office location  
Västeråsen  
Brinellgatan 4  
SE-504 62 Borås  
SWEDEN

Phone / Fax / E-mail  
+46 10 516 50 00  
+46 33 13 55 02  
info@sp.se

Laboratories are accredited by the Swedish Board for Accreditation and Conformity Assessment (SWEDAC) under the terms of Swedish legislation. This report may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.