



PRODUCT INFORMATION UPDATE - September 2008

Power Wheelchairs - Mirage

Please find below details of the latest product changes for our Power chair range.

Item	Description of change	Picture
MIRAGE	<p>Due to continual improvement activities, the circuit breaker on the Mirage has been changed from a spade type connector to a bolt type connector which will ensure good connectivity.</p>	
	<p>The serial number of the first Mirage with the new circuit breaker is 060813905024. Production date is 12.08.2008</p>	

If you require any additional information regarding any of the above changes, please contact your local Invacare

TECHNICAL SERVICES

TEL: 01656 647327

FAX: 01656 753299

EMAIL: uk@invacare.com